



FERRARA RING™

Intrastromal Ferrara Ring

<i>Brand name</i>	FERRARA RING™
<i>Reference</i>	AFR/FR (5 mm optical zone) and AFR6/FR6 (6 mm optical zone).
<i>Definition</i>	Intrastromal Corneal Implants.
<i>Arc</i>	90°, 100°, 110°, 120°, 130°, 140°, 150°, 160°, 170°, 180°, 190°, 200°, 210°, 300°, 320°, 340°
<i>Thickness</i>	0.15, 0.20, 0.25, 0.30, 0.35 mm
<i>Indications</i>	<p>The correct indication of Ferrara Ring implants requires a detailed evaluation of the topographic conditions and pachymetry of the cornea, in addition to a complete visual examination.</p> <ul style="list-style-type: none"> • Progressive keratoconus. • Keratoconus with intolerance to contact lenses. • Corneal deformity induced by contact lenses (Harstein's syndrome). • Post-penetrating keratoplasty astigmatism. • Post-Lasik corneal ectasias. • Post-radial keratotomy astigmatism. • Pellucid marginal degeneration.
<i>Product description</i>	<p>The FERRARA RING™ consists of a semicircular segment with an arc of multiple lengths and a fixed triangular section: AFR (0.60 mm base) and AFR6 (0.80 mm base).</p> <p>Single use implant. Each segment has a 0.20 mm hole in their apex, to facilitate its implantation.</p>
<i>Material</i>	Polymethyl Methacrylate (PMMA) with natural blue light filter.
<i>Sterilization method</i>	Ethylene Oxide (ETO).
<i>Supply</i>	Individual sterile segment, not pyrogenic
<i>Shelf-life</i>	4.5 years

REFERENCE TABLE FOR ORDERS

<i>Optical zone</i>	AFR= 5 mm optical zone	Reference: AFRXXYYY XX= Ring thickness AFR6XXYYY YYY= Ring arc length
	AFR6= 6 mm optical zone	
<i>Thickness</i>	0.15 mm, 0.20 mm, 0.25 mm, 0.30 mm, 0.35 mm.	Example: AFR15160: AFR620150: 5 mm optical zone 6 mm optical zone Thickness 0.15 mm Thickness 0.20 mm Arc 160° Arc 150°
<i>Arc length</i>	90°, 100°, 110°, 120°, 130°, 140°, 150°, 160°, 170°, 180°, 190°, 200°, 210°, 300°, 320°, 340°.	

CE 0318



V-04-01-AAB-IMP REV.00